

Final Data Validation Report

USACE Fort Wingate Depot Activity New Mexico

Project No: Eco-18-1237

SDG #23D225 Analytical Data Package

Publication Date: 09-07-2023

By
ZIBA HOSSEINI
14343 Peach Hill Rd.
Moorpark, CA, 93021

TABLE OF CONTENTS

EXECUTIVE SUMMARY	4
1.0 INTRODUCTION	
1.1 Objectives and Scope of the Data Validation	6
1.2 Organization of the Report	6
2.0 DATA REVIEW AND VALIDATION	
2.1 Data Reporting	7
2.2 Data Evaluation	8
2.2.1 Holding Times	15
2.2.2 Laboratory and Field Blanks	16
3.0 QUALITY ASSURANCE OBJECTIVES	
3.1 Qualitative QA Objectives	17
3.1.1 Comparability	17
3.1.2 Representativeness	17
3.2 Quantitative QA Objectives	17
3.2.1 Precision	18
3.2.2 Accuracy	18
3.2.3 Completeness	19
4.0 DATA VALIDATION	
4.0 Results of Data Validation	20
4.1. VOCs by EPA Method 8260C	21
4.2. Semivolatiles + Extra analytes by EPA Method 8270D	25
4.3. Organochlorine Pesticides by EPA Method 8081B	30
4.4. Polychlorinated Biphenyls by EPA Method 8082A	32
4.5. Chlorinated Herbicides EPA Method 8151A	34
4.6. Nitroaromatics & Nitramines by EPA Method SW846 3535A/8330B	35
4.7. Nitroglycerine and PETN by EPA Method 8332	37
4.8. Total Petroleum Hydrocarbons (GROs) By EPA Method 8015D	38
4.9. Total Petroleum Hydrocarbon (DROs) by EPA Method 8015D	40

4.10.	Perchlorate by EPA Method 6850	41
4.11.	Total and Dissolved Metals by ICP-MS (EPA Method 6020A)	42
4.12.	Mercury & Dissolved Mercury by EPA Method 7470A	45
4.13.	Anions by SW9056A	46
5.0	CONCLUSION	49
6.0	REFERENCES	49

EXECUTIVE SUMMARY

This data validation report presents the evaluation and validation of the analytical data for samples collected in April 2023 as part of water monitoring, Fort Wingate Depot Activity, New Mexico (NM). EMAX Laboratories in Torrance, California performed the chemical analysis of these samples. The United States Army Corps of Engineers and the State of California have certified EMAX Laboratories to perform the analysis described within this project, (QAPP, Eco & Associate, Inc. Project number Eco-18-1237, April 2019).

A total of twelve (12) water samples were collected on 04-19-23. EMAX Laboratories received the samples on 04-20-23. Data was delivered in one package as stage 2b and stage 3 deliverable. Ten percent of the data was subjected to validation equivalent to stage 3 deliverable. Raw data for all samples were submitted for the requested analytical methods. One sample from this sample delivery group, TMW50042023(Lab ID#D225-03) was designated as stage 3 deliverable. Raw data for this sample was compared to the reported summary tables for each method and went through comprehensive data validation review. Sample TMW63042023 (Lab ID#D225-08) was designated to be spiked as MS/MSD on the chain of custody. Raw data for this sample together with method blank and LCS/LCSD for each method were also cross checked with the corresponding summary table results.

Stage 2b data validation examined quality assurance/quality control (QA/QC) elements such as holding time, (both extraction and analysis), extraction logs, instrument injection logs, method blank results, QC summary results and recoveries, LODs/LOQs, summaries of initial and continuing calibrations and completeness of results for the following requested EPA methods of analysis:

EPA Method 3050B/8260C: Volatile Organics by GC-MS (12 samples)

EPA Method 3520C/8270D: Semi-Volatile Organics by GC/MS (8 samples)

EPA Method SW846 3535A/8330B: Nitroaromatics and Nitramines (9 samples)

EPA Method SW8332: Nitroglycerine and PETN (9 samples)

EPA Method 8081B: Organochlorine Pesticides (7 samples)

EPA Method 8082A: Polychlorinated Biphenyls; PCBs (4 sample)

EPA Method 8151A: Chlorinated herbicides (4 sample)

EPA Method 8015D: Total Petroleum Hydrocarbons (GROs) (6 samples)

EPA Method 8015D: Total Petroleum Hydrocarbons; extractable (DROs) (5 samples)

EPA Method 6850: Perchlorate (11 samples)

EPA Method SW6020A: Dissolved and total Metals by ICP-MS (11 samples)

EPA Method 7470A: Mercury & Dissolved Mercury (11 samples)

EPA Method SW9056A: Nitrate-N, Nitrite, Orthophosphate, Bromide, Chloride, Fluoride and Sulfate (10 samples)

The analytical results, QC results, initial calibration summary table and initial calibration verification (ICV) data were comprehensively compared with the corresponding raw data and chromatograms presented for stage 3 data validation.

All the requested samples were analyzed for each of the components listed in the corresponding EPA Methods (QAPP; final version, Eco & Associate, Inc. April 2019). The evaluation indicated that all the analytical work was performed as requested on the chain of custody. The extraction and analytical holding times were met for each method and all the related samples. The deviations, if any, are discussed in Section 4.0 for each method.

The SDG # 23D225 analytical data evaluated in this data validation report has met the data quality and usability requirement as defined in the data quality objectives. While very few analytical QC exceedances were observed, it was not significant for any data qualifiers. Overall data is of acceptable quality and considered usable for its intended purpose.

1.0 INTRODUCTION

This report presents the evaluation and validation of analytical data for water samples collected as a part of water monitoring at Fort Wingate, New Mexico (NM).

1.1 Objectives and Scope of Data

The main objective of this report is to assess the acceptability of the data generated by the designated laboratory. The data validation was performed according to the analytical requirements of the method in the *Quality Assurance Project Plan, final Draft, USACE Fort Wingate New Mexico*, (Project No: Eco-18-1237, April 2019), *USEPA Analytical Operations/Data Quality Center (AOC) National Functional Guidelines for Organic Data Review* (USEPA, January 2017), *National Functional Guidelines for Inorganic Data Review* (USEPA, September 2016), *US DoD General Data Validation Guideline*, February 2018, *EM 200-1-10 Guidance for Evaluating Performance-based Chemical Data, US Army Corps of Engineers (USACE)*, June 2005 and *DoD Quality System Manual DoD QSM 5.3, 2019*. The Approved site-specific Quality Assurance Project Plan (ECO QAPP) has the highest hierarchy.

1.2 Organization of the Report

Section 2.0 describes the components of the data review. Section 3.0 provides the qualitative quality assurance objectives. Section 4.0 summarizes the findings and conclusions of the data validation.

2.0 DATA REVIEW AND VALIDATION

Data validation is a systematic method for reviewing and qualifying the presented analytical data for their intended use. The objective of this data validation report is to identify any unacceptable or faulty measurements, as reported by the laboratory.

EMAX Laboratories in Torrance, California performed the chemical analysis of the samples. Army Corps of Engineers and the State of California has certified this laboratory for performing the analysis described within this report.

A total of twelve (12) water samples were collected on 04-19-23. EMAX Laboratories received the samples on 04-20-23.

2.1 Data Reporting

The data was delivered in one package as stage 2b and stage 3 deliverables. 10% of the data was subjected to validation to the equivalent of stage 3.

EMAX Laboratories provided the following information in one data package:

- Sample identification number;
- Date of sample collection;
- Sample matrix type;
- Analysis method;
- Target lists and results of analysis;
- Limit of Detection (LOD);
- Limit of Quantitation (LOQ);
- Laboratory qualifiers and qualifier definitions;
- Copies of sample logs and chain-of-custody logs;
- Sample preparation logs (with the sample extraction dates);
- Sample Analysis logs (Instrument injection logs with sample analysis dates);
- Results and percent recoveries of Matrix Spike Samples (MS/MSD), if submitted
- Results and percent recoveries of Lab Control Samples (LCS/LCSD)

- Summary of initial calibration, initial calibration verification (ICV) and continuing calibration verification (CCV) standards;
- Case narrative for each method;
- Raw data for all the initial calibrations, initial calibration verifications, continuing calibrations, Tune check standards (where applicable), internal standard responses and chromatograms for the sample/samples at Stage 3 deliverable and related QC samples.

Data validation was performed by initial review of the analytical reports and QA/QC results and recoveries using summary tables. Next, selected analytical reports including QA/QC information was cross checked with raw data. The analysis and extraction sequence logs for each method were examined. Overall review assessed the effects of QA/QC results on the data usability. The review included such parameters as holding times, LODs/LOQs, initial and continuing calibration method requirements, surrogate recoveries, MS/MSD and lab control samples (LCS/LCSD) results and percent recoveries for accuracy and precision.

Stage 3 review compared the reported analytical results with those obtained from the raw data. Raw data for each analytical method requested on the chain of custody were submitted for all samples. One field sample TMW50042023 (EMAX ID #D225-03) from this sample delivery group was designated for stage 3 data review. Raw data for this sample was evaluated comprehensively. Sample TMW63042023(EMAX ID #D225-08) was designated to be spiked as MS/MSD on the chain of custody. Raw data for method blank, LCS/LCSD and MS/MSD was reviewed in detail. Calculations and corresponding equations, as well as analyte identification were randomly checked and verified.

2.2 Data Evaluation

The following parameters were evaluated in the preliminary data review:

- Analysis performed and sample identifications were verified to be in accordance with the information provided on the chain-of-custody (COC);

- Technical holding times were confirmed for all samples with regard to the requested method of analysis (collection to extraction and extraction to analysis);
- Limit of quantitation (LOQ) for each analyte reported were compared with the project measurement objectives;
- Initial calibration and initial calibration verification standards were evaluated;
- Continuing calibration standards were evaluated
- Trip blank results (Method 8260C and TPH by purge & trap only) were evaluated;
- MS/MSD results and recoveries were evaluated; only for total metals
- LCS/LCSD results and recoveries were evaluated; and
- Method blank results as well as surrogate recoveries, internal standards, instrument performance check compounds (for GC-MS) and DDT/Endrin breakdown (Method 8081B) were evaluated.

The following is a list of sample identifications and corresponding laboratory sample identification numbers:

Site Name: Fort Wingate, New Mexico				
SDG#23D225			Matrix: Water	
Field/Client ID	Lab ID	Date collected	Validation stage	Requested Methods of Analysis
TMW30042023	23D225-01	04-19-23	S3VM	VOCs by SW5030B/8260C, Semivolatiles + APP9 Organochlorine Pesticides Nitroaromatics and Nitramines Nitroglycerine & PETN Dissolved Mercury Mercury Dissolved Metals by ICP-MS Total Metals by ICP-MS Perchlorate by 6850 Anions by IC
TMW26042023	23D225-02	04-19-23	S3VM	VOCs by SW5030B/8260C, Organochlorine Pesticides Nitroaromatics and Nitramines Nitroglycerine & PETN Mercury & Dissolved Mercury Dissolved Metals by ICP-MS Total Metals by ICP-MS Perchlorate by 6850 Anions by IC
TMW500420223	23D225-03	04-19-23	S3VM	VOCs by SW5030B/8260C, Semivolatiles + APP9 Organochlorine Pesticides Nitroaromatics and Nitramines Nitroglycerine & PETN Dissolved Mercury Mercury Dissolved Metals by ICP-MS Total Metals by ICP-MS Chlorinated Herbicides Polychlorinated Biphenyls Perchlorate by 6850 TPH Gasoline; TPH as DRO Anions by IC
SMW010420223	23D225-04	04-19-23	S3VM	VOCs by SW5030B/8260C, Semivolatiles + APP9 Nitroaromatics and Nitramines Nitroglycerine & PETN Mercury & Dissolved Mercury Dissolved Metals by ICP-MS Total Metals by ICP-MS Perchlorate by 6850 Anions by IC

Site Name: Fort Wingate, New Mexico				
SDG#23D225			Matrix: Water	
Field/Client ID	Lab ID	Date collected	Validation	Requested Methods of Analysis
SMW01042023D	23D225-05	04-19-23	S3VM	VOCs by SW5030B/8260C, Semivolatiles + APP9 Nitroaromatics and Nitramines Nitroglycerine & PETN Mercury & Dissolved Mercury Dissolved Metals by ICP-MS Total Metals by ICP-MS Perchlorate by 6850 Anions by IC
TMW31S042023	23D225-06	04-19-23	S3VM	VOCs by SW5030B/8260C, Semivolatiles + APP9 Organochlorine Pesticides Nitroaromatics and Nitramines Nitroglycerine & PETN Dissolved Mercury Mercury Dissolved Metals by ICP-MS Total Metals by ICP-MS Perchlorate by 6850 Organochlorine Pesticides Anions by IC
BGMW11042023	23D225-07	04-19-23	S3VM	VOCs by SW5030B/8260C, Semivolatiles + APP9 Organochlorine Pesticides Nitroaromatics and Nitramines Nitroglycerine & PETN Dissolved Mercury Mercury Dissolved Metals by ICP-MS Total Metals by ICP-MS Perchlorate by 6850 TPH Gasoline; TPH as DRO Chlorinated Herbicides Polychlorinated Biphenyls Anions by IC
TMW63042023	23D225-08	04-19-23	S3VM	VOCs by SW5030B/8260C, Semivolatiles + APP9 Organochlorine Pesticides Nitroaromatics and Nitramines Nitroglycerine & PETN Dissolved Mercury Mercury Dissolved Metals by ICP-MS Total Metals by ICP-MS Perchlorate by 6850 TPH Gasoline; TPH as DRO Chlorinated Herbicides Polychlorinated Biphenyls Anions by IC

Site Name: Fort Wingate, New Mexico				
SDG#23D225			Matrix: Water	
Field/Client ID	Lab ID	Date collected	Validation stage	Requested Methods of Analysis
TMW08042023	23D225-09	04-19-23	S3VM	VOCs by SW5030B/8260C, Nitroaromatics and Nitramines Nitroglycerine & PETN Dissolved Mercury Mercury Dissolved Metals by ICP-MS Total Metals by ICP-MS Perchlorate by 6850 TPH Gasoline; TPH as DRO Anions by IC
QC19042023EB2	23D225-10	04-19-23	S3VM	VOCs by SW5030B/8260C, Semivolatiles + APP9 Organochlorine Pesticides Nitroaromatics and Nitramines Nitroglycerine & PETN Dissolved Mercury Mercury Dissolved Metals by ICP-MS Total Metals by ICP-MS Perchlorate by 6850 TPH Gasoline; TPH as DRO Chlorinated Herbicides Polychlorinated Biphenyls Anions by IC
TMW27042023	23D225-11	04-19-23	S3VM	VOCs by SW5030B/8260C, Semivolatiles + APP9 Nitroaromatics and Nitramines Nitroglycerine & PETN Mercury Dissolved Mercury Dissolved Metals by ICP-MS Total Metals by ICP-MS Perchlorate by 6850

Site Name: Fort Wingate, New Mexico				
SDG#23D225			Matrix: Water	
Field/Client ID	Lab ID	Date collected	Validation stage	Requested Methods of Analysis
QC19042023TB03	23D225-12	04-17-23	S3VM	VOCs by SW5030B/8260C, TPH Gasoline;
TMW63042023MS	23D225-08M	04-19-23	S3VM	VOCs by SW5030B/8260C, Semivolatiles + APP9 Organochlorine Pesticides Nitroaromatics and Nitramines Nitroglycerine & PETN Dissolved Mercury Mercury Dissolved Metals by ICP-MS Total Metals by ICP-MS Perchlorate by 6850 TPH Gasoline; TPH as DRO Chlorinated Herbicides Polychlorinated Biphenyls
TMW63042023MSD	23D225-08S	04-19-23	S3VM	VOCs by SW5030B/8260C, Semivolatiles + APP9 Organochlorine Pesticides Nitroaromatics and Nitramines Nitroglycerine & PETN Dissolved Mercury Mercury Dissolved Metals by ICP-MS Total Metals by ICP-MS Perchlorate by 6850 TPH Gasoline; TPH as DRO Chlorinated Herbicides Polychlorinated Biphenyls

TABLE 2-1
Summary of Analytical Parameters
USACE Wingate, New Mexico

Table 2-1 below shows the specified analysis for constituents in the water samples, the corresponding Environmental Protection Agency (EPA) analytical method, and the corresponding limit of quantitation (LOQ), of groups of constituents.

MATRIX	CONSTITUENT	EPA METHOD	LOQ
Water	Volatile Organic Compounds list	SW5030B/8260C	1,2 & 20 µg/L
	Semi Volatile Organic Compound List	SW3520C /8270D	10&20µg/L, (Benzidine=40µg/L)
	Nitroaromatics & Nitramines	SW3535A/8330B	0.4µg/L
	Nitroglycerine & PETN	SW3535A/8332B	0.4µg/L
	Chlorinated Herbicides	SW8151A	1µg/L, (MCPA=40µg/L)
	Organochlorine Pesticides	SW8081B	0.1µg/L Methoxychlor =1.0µg/L Toxaphene =2.0µg/L
	Polychlorinated Biphenyls (PCBs)	SW8082A	1µg/L
	Total Petroleum Hydrocarbons (GROs)	SW8015D Purge & Trap	100µg/L
	Total Petroleum Hydrocarbons (DROs)	SW8015D Extractable	0.5mg/L
	Dissolved & Total Metals By ICP-MS	SW6020A	0.5µg/L, 1µg/L, 20µg/L, 100µg/L, 200µg/L
	Dissolved Mercury/Mercury	SW7470A	0.5µg/L
	Anions by IC	SW9056A	0.1mg/L; 0.2mg/L; 0.5mg/L
	Perchlorate	SW6850	0.5µg/L

2.2.1 Sample Receipt

Documentations and recordings regarding status of each sample and cooler temperatures upon receipt in the laboratory were reviewed. Samples were received in fifteen ice preserved coolers.

2.2.2 Holding Times

Technical holding times are defined as the maximum time allowed between sample collection, extraction, and analysis. Collection to extraction and extraction-to-analysis (40-day) was within the holding time requirement for semi-volatile organic methods. Extraction-to-analysis was within the method's holding time requirement with metals and inorganic methods. Table 2-2 presents the summary of holding time requirements with qualifications if applied.

TABLE 2-2
Summary of Analytical Methods and Holding Time Requirements
USACE Wingate, New Mexico

ANALYSIS Method	MATRIX	HOLDING TIME REQUIREMENT	DATA QUALIFIED AS "J"
EPA Method 5030B/8260C	Water	14days to analysis (7days if not acid preserved)	None. Holding times were met
Semi Volatile Organic Target List 3520C/8270D/8270SIM	Water	Collection to extraction: 7 days Extraction to analysis: 40 days	None. Holding times were met
Nitroaromatics and Nitramines	Water	Collection to extraction: 7 days Extraction to analysis: 40 days	None. Holding times were met
Nitroglycerine and PETN	Water	Collection to extraction: 7 days Extraction to analysis: 40 days	None. Holding times were met
Chlorinated Herbicides	Water	Collection to extraction: 7 days Extraction to analysis: 40 days	None. Holding times were met
Organochlorine Pesticides	Water	Collection to extraction: 7 days Extraction to analysis: 40 days	None. Holding times were met
Polychlorinated Biphenyls (PCBs)	Water	Collection to extraction: 7 days Extraction to analysis: 40 days	None. Holding times were met
Total Petroleum Hydrocarbons (GROs)	Water	14days to analysis (7days if not acid preserved)	None. Holding times were met
Total Petroleum Hydrocarbons (DROs)	Water	Collection to extraction: 7 days Extraction to analysis: 40 days	None. Holding times were met
Perchlorate	Water	Collection to Analysis: 28 days	None. Holding times were met
Dissolved and Total Metals	water	Analysis within 6 Months	None. Holding times were met
Anions by IC	Water	Analysis 48 hours from collection for Nitrate, Nitrite & Orthophosphate and 28-days for Bromide, Chloride, Fluoride and Sulfate	Holding times were met for all
Mercury & Dissolved Mercury	Water	Collection to Analysis: 28 days	None. Holding times were met

2.2.3 Laboratory and Field Blanks

The objective of laboratory and field blanks is to determine the presence and extent of contamination resulting from laboratory or field activities. Blanks reported here included method and/or extraction blanks and trip blanks (VOCs and Gasoline only). The result of analysis of method blank is discussed in Section 4.0 for each method. All samples were transported in fifteen ice preserved coolers and were stored in a refrigerator upon arrival to the laboratory. The cooler's temperature was reported as low as 1.5°C and as high as 5.9°C upon arrival. All samples were received intact and in good condition.

3.0 QUALITY ASSURANCE OBJECTIVES

Quality assurance (QA) objectives define analytical parameters that validate the conclusions drawn from the results. Quality assurance was assessed through the following means: precision, accuracy, representativeness, completeness, and comparability (PARCC).

3.1 Qualitative QA Objectives

Qualitative aspects of QA for analytical data are characterized by completeness and representativeness.

3.1.1 Comparability

Comparability defines the level of confidence with which one data set can be compared with another. Comparability is related to accuracy and precision. It is also a measure of the data's reliability. All units for comparability are in accordance with standard procedures so that the results could be compared with other laboratories if necessary.

3.1.2 Representativeness

Representativeness is a quantity, which presents whether the results of analysis accurately portray the actual site conditions. Representativeness is a qualitative parameter, which signifies the extent of accuracy and precision, to which the data represent a characteristic population, parameter variations at a sampling point, process condition, or environmental conditions. The sampling procedures described within the approved QAPP (Eco & Associate, Inc., April 2019) are designed to provide samples representative of the site conditions.

3.2 Quantitative QA Objectives

Quantitative QA Objectives for analytical data are defined as precision, accuracy, completeness, and method quantitation limits. These quantitative parameters are established in order to monitor the overall quality of analytical data produced by the laboratory. The laboratory performing the analytical methods specified in Table 2-1, and the case narratives, which is included in the data package from the laboratory, ensures the quality of the analytical data.

3.2.1 Precision

Precision is a measure of the closeness with which multiple analyses of a given sample agree with each other. It describes the agreement between two or more measurements that have been made in exactly the same way. Precision is measured through matrix spike/matrix spike duplicate samples, laboratory control sample/ laboratory control sample duplicate and sample/sample duplicate analysis. In the latter case, the sample with positive results can be used for this purpose. The relative percent difference (RPD) is calculated as a means of quantifying precision. The following equation is used for this purpose:

$$RPD = \frac{R_1 - R_2}{(R_1 + R_2)/2} \times 100$$

Where:

RPD = Relative percent difference

R₁ = Result of the first duplicate or measured sample concentration

R₂ = Result of the second duplicate or known sample or duplicate concentration

When analytes are present at concentrations below or near the quantitation limit, precision is measured, using MS/MSD, and/or LCS/LCSD results.

Precision results are discussed in Section 4.0 of this report.

3.2.2 Accuracy

Accuracy indicates the closeness of the measurement to its true or accepted value. Accuracy measures agreement between a result and its true value. Accuracy is measured through laboratory control sample analysis and surrogate recoveries. Method-specific QA objectives for precision and accuracy were based on the quality control limits developed by the laboratory for the analytical methods, specified in Table 2-1. These procedures may affect the accuracy of the data presented. Additionally, initial and continuing calibrations were used to verify that the analytical instrument accurately measured the compound concentrations. Calculations were independently verified for the responses and percent differences (%Ds).

3.2.3 Completeness

Completeness is defined as the percentage of total measurements, which are judged to be valid. The completeness objective is to obtain a sufficient amount of valid data to enable the goals and objectives of the project to be achieved.

Completeness is quantified by computing the fraction of reports, which remained valid after the sampling procedures were reviewed and the results conformed to QA/QC protocols. The following equation was used to calculate completeness:

$$\text{Completeness} = \frac{\text{No. of valid field samples collected and analyzed}}{\text{No. of valid field samples reported}} \times 100$$

Completeness (EPA Method 5030B/8260C: VOCs) = $12/12 \times 100 = 100\%$

Completeness (EPA Method 3520B/8270D: SVOCs) = $8/8 \times 100 = 100\%$

Completeness (EPA Method 3535A/8330B: Explosives) = $9/9 \times 100 = 100\%$

Completeness (EPA Method 8332: Nitroglycerine & PETN) = $9/9 \times 100 = 100\%$

Completeness (EPA Method 8081B: Organochlorine pesticides) = $7/7 \times 100 = 100\%$

Completeness (EPA Method 8082A: Polychlorinated Biphenyls) = $4/4 \times 100 = 100\%$

Completeness (EPA Method 8151B: Chlorinated Herbicides) = $4/4 \times 100 = 100\%$

Completeness (EPA Method 8015G: Petroleum Hydrocarbons; GRO) = $6/6 \times 100 = 100\%$

Completeness (EPA Method 8015D: Petroleum Hydrocarbons; DRO) = $5/5 \times 100 = 100\%$

Completeness (EPA Method 6850: Perchlorate) = $11/11 \times 100 = 100\%$

Completeness (EPA Method 7470A: Mercury & Dissolved Mercury) = $11/11 \times 100 = 100\%$

Completeness (EPA Method 6020A: Dissolved and Total Metals) = $11/11 \times 100 = 100\%$

Completeness (EPA Method SW9056A: Nitrate-N, Nitrite, Orthophosphate, Bromide, Chloride, Fluoride and Sulfate) = $10/10 \times 100 = 100\%$

Completeness is affected by anything that reduces the number of samples analyzed (such as a sample loss during transport or extraction), as well as acceptance or non-acceptance of analytical results.

4.0 DATA VALIDATION

This data review covers twelve water samples listed on page 10 including dilutions and reanalysis if applicable. The analyses were according to the following EPA Methods:

EPA Method **5030B/8260C** for VOCs by GC/MS

EPA Method **3520C/8270D** for SVOCs by GC/MS

EPA Method **8081B** for Organochlorine pesticides by GC/ECD

EPA Method **8082A** for Polychlorinated Biphenyls

EPA Method **8151B** for Chlorinated Herbicides

EPA Method **8015D (GROs)**, Total Petroleum Hydrocarbons by GC/FID

EPA Method **8015D (DROs)**, Total Petroleum Hydrocarbons by GC/FID

EPA Method **3535A/8330B** for Nitroaromatics and Nitramine by HPLC/UV

EPA Method **8332** for Nitroglycerine and PETN by UHPLC/UV

EPA Method **6850** for Perchlorate by HPLC/MS

EPA Method **6020A** for Dissolved and total metals by ICP/MS

Method **7470A** for Mercury & Dissolved Mercury by Cold Vapor

EPA Method **9056A** for Anions by IC

This review follows *Quality Assurance Project Plan, final Draft, USACE Fort Wingate Depot Activity*, McKinley County, New Mexico; Project # Eco-18-1237 April 2019, EM 200-1-10 *Guidance for Evaluating Performance-based Chemical Data; US Army Corps of Engineers (USACE). June 2005*, and USEPA Analytical Operations/Data Quality Center (AOC) *National Functional Guidelines for Organic Data Review* (USEPA, January 2017); DoD QSM 5.3, 2019 and *National Functional Guidelines for Inorganic Data Review* (USEPA, September 2016). The Approved site-specific Quality Assurance Project Plan has the highest hierarchy. The following subsections correlate to the above guidelines.

The followings are definitions of the data qualifiers:

- U Indicates the analyses was analyzed for but not detected at or above Limit of Detection (LOD).
- J Indicates an estimated value with an unknown bias.

- UJ Indicates the analyte was analyzed for but not detected and reported less than LOD. However, the numerical value is approximate.
- J⁺ The result was estimated value and may be biased high.
- J The result was estimated value and may be biased low.

- X The sample results (including non-detects) were affected by serious deficiencies in the ability to analyze the sample and to meet published method and project quality criteria. The presence or absence of the analyte cannot be confirmed by the data provided. Acceptance or rejection of the data should be decided by the project team, but exclusion of the data is recommended

The following Reason codes were applied in the report:

- M3 MS/MSD and/or LCS/LCSD percent recovery infraction with low bias
- M4 MS/MSD or duplicate precision infraction
- S1 Surrogate percent recovery infraction with high bias
- S2 Surrogate percent recovery infraction with low bias
- R4 Result exceeds calibration range
- B6 Trip blank infraction (qualified detect)
- B7 Field blank infraction (qualified detect)
- B8 Equipment blank infraction (qualified detect)
- D1 Field duplicate precision infraction

4.1. VOC (EPA Method 5030B/8260C)

4.1.1. Technical Holding Times

Holding time requirement was met for all twelve samples. Water samples were collected on 04-19-23. Samples were analyzed on 04-28-23 and 04-29-23 within method's requirement for holding time. (Water samples were preserved with hydrochloric acid).

The chain-of-custody was reviewed for documentation of sample information and method of analysis.

Qualification notations, if any, will be summarized in result section; section 4.1.7.

4.1.2. Tuning criteria

Performance of the instrument was checked by injection of a single component tune check standard (BFB: Bromofluorobenzene) prior to initial calibration on 04-26-23 and at the beginning of each analysis shift on 04-28-23 and 04-29-23. It passed all the method assigned criteria.

4.1.3. Initial Calibration

Samples were analyzed with reference to one set of initial calibration using GC/MSD. Initial calibration curve was generated on 04-26-23. A multi-level calibration standard ranging from 0.3µg/L to 100µg/L was used for this purpose. Internal standard curve type was used for initial calibration and all following analysis. Minimum response factor for system performance check compounds (SPCCs) were within the method acceptable limits (Table 4.1.3.1). Response factors at each level were randomly recalculated and all agreed with the response factors submitted in the initial calibration summary table.

Minimum average response factors for the system performance check compounds (SPCCs) were recognized according to the following table:

Table 4.1.3.1: System Performance Check Compounds (Initial calibration)

System Performance check compounds (SPCCs)	Minimum average response factor (requirement)	Average Response factor 04-26-23
Chloromethane	≥ 0.10	√
1,1-Dichloroethane	≥ 0.20	√
Bromoform	≥ 0.10	√
Chlorobenzene	≥ 0.50	√
1,1,2,2-Tetrachloroethane	≥ 0.30	√

√ denotes passing method acceptance limits

Average response factor curve fit was used mainly through the initial calibration. All target compounds met the maximum 15% RSD limit.

Minimum average response factors for all target compounds were within method's recommended values, except for: Acetone (0.040) and 2-Butanone (0.013). However, recoveries were within the requirement of 70-130%. Calibration check compounds (CCCs) met the acceptance criteria for %RSD among the response factors calculated for each level. Table 4.1.3.2

lists the CCCs with method requirement limits for %RSD among response factors for initial calibration.

Table 4.1.3.2 Calibration Check Compounds (CCCs) Initial Calibration

Calibration Check Compounds (CCCs)	Response Factors %RSD (Method limit)	Response Factors %RSD 04-26-23
Vinyl chloride	≤ 20%	√
1,1-Dichloroethene	≤ 20%	√
Chloroform	≤ 20%	√
1,2-Dichloropropane	≤ 20%	√
Toluene	≤ 20%	√
Ethyl benzene	≤ 20%	√

√ denotes passing method acceptance limits

4.1.4. Initial Calibration Verification and Continuing Calibration

Initial calibration was verified by a second source standard on 04-26-23. Percent difference (%D) and/or drift between initial calibration RRFs (average response factors) and the initial calibration verification response factors (RFs) were less than or equal to 20% for all target compounds.

Continuing calibration check standard was analyzed at the beginning and end of each analysis shift on 04-28-22 and 08-29-23. Prior to continuing calibration standard, **instrument performance check standard** (BFB tune check) was carried out. It passed all the method tuning criteria.

Minimum average response factors for the system performance check compounds (SPCCs) were all within the method limits according to the following table:

Table 4.1.4.1: System Performance Check Compounds (Daily calibration)

System Performance Check compounds (SPCCs)	Minimum response factor (Method limits)	Continuing cal. Response factors 04-28-23 (I &II)	Continuing cal. Response factors 04-29-23 (I &II)
Chloromethane	≥ 0.10	√	√
1,1-Dichloroethane	≥ 0.20	√	√
Chlorobenzene	≥ 0.50	√	√
Bromoform	≥ 0.10	√	√
1,1,2,2-Tetrachloroethane	≥ 0.30	√	√

√ denotes passing method acceptance limits

Calculated percent differences (%D) between initial calibration RRFs (average response factors) and the continuing calibration response factors (RFs) were less than or equal to 20% for all the Calibration Check Compounds (CCCs) and less than or equal to 20% for nearly all other target analytes. Area counts for all internal standards were within ± 50 percent of the same level in the initial calibration. The calculated % difference between RFs from continuing calibration and average response factors from initial calibration is summarized in Table 4.1.4.2 for continuing calibration reports presented with the data package.

Table 4.1.4.2 Calibration Check Compounds (CCCs): Continuing Calibration

Calibration Check Compounds (CCCs)	%Deviation From Initial calibration (Acceptance Limit)	%Deviation from Initial calibration (04-28-23) I&II	%Deviation from Initial calibration (04-29-23) I&II
Vinyl chloride	$\leq 20\%$	√	√
1,1-Dichloroethene	$\leq 20\%$	√	√
Chloroform	$\leq 20\%$	√	√
1,2-Dichloropropane	$\leq 20\%$	√	√
Toluene	$\leq 20\%$	√	√
Ethyl benzene	$\leq 20\%$	√	√

√ denotes passing method acceptance limits

Deviation from the initial calibration was less than 20 percent for the rest of target list, except for one target analyte listed in the table below for each opening and closing daily standards. This should not affect the data quality.

Target analyte	%Deviation From Initial calibration (Acceptance Limit)	%Deviation from Initial calibration (04-29-23)	
		I	II
Carbon disulfide	$\leq 20\%$	√	35.1%*

*Outside control limit

4.1.5. Quality Control samples reported consisted of two method blanks, two sets of LCS/LCSD and MS/MSD. Sample TMW63042023 was designated to be analyzed as MS/MSD. The full list of target compounds was spiked and reported for LCS/LCSD and MS/MSD. Percent

recoveries and percent RPDs for QC samples reported, were within the project acceptance limits for the entire compound list.

The results, percent recoveries and RPDs were recalculated randomly and all agreed with the reported QC summary table. Method blank presented with the data package, analyzed with samples did not show presence of any target compounds.

Surrogate recoveries were all within the method's acceptable limits.

4.1.6. Field duplicate sample and its associated sample: Field sample SMW01042023 was identified as field duplicate of sample SMW01042023D. No VOCs was detected in any of field sample or associated field duplicate sample.

4.1.7. Raw data was submitted for all samples. Sample TMW50042023 (Lab ID# 23D225-03) was designated to be reviewed as stage 3 data deliverable. Raw data for this sample with all related QC samples was reviewed for stage 3 data validation. The results calculated from the raw data agreed with all the results reported in data summary reports.

4.2. SVOC (EPA Method 3520C/8270D)

4.2.1. Technical Holding Times

Holding time requirement was met for all samples. Eight water samples were collected on 04-19-23, extracted on 04-26-23 and were analyzed on 04-28-23 within required holding time.

The chain-of-custody was reviewed for documentation of sample information and method of analysis.

Qualification notations, if any, will be summarized in result section; section 4.2.7.

4.2.2. Tuning criteria

Performance of the instrument was checked by injection of a tune check standard (DFTPP: Decafluorotriphenylphosphine) prior to each set of initial calibration on 03-07-22, 03-31-23 and at the beginning of each analysis shift on 04-28-23. It passed all the method assigned criteria. Tailing of Pentachlorophenol and Benzidine was within acceptance limit (less than 2%) and breakdown of DDT was also within methods acceptance limit of less than 20%.

4.2.3. Initial Calibration

Samples were analyzed with reference to one set of initial calibration using GC/MSD. Due to long list of analytes used for this method, three separate lists of compounds were grouped together and initial calibration was generated separately for each group.

Initial calibration curves were generated on 03-07-22 and 03-31-23. A multi-level calibration standard ranging from 4mg/L to 50mg/L was used for this purpose. Internal standard curve type was used for initial calibration and all following analysis. Minimum response factor for system performance check compounds (SPCCs) were within the method acceptable limits (Table 4.2.3.1). Response factors at each level were randomly recalculated and all agreed with the response factors submitted in the initial calibration summary table.

Minimum average response factors for the system performance check compounds (SPCCs) were recognized according to the following table.

Table 4.2.3.1: System Performance Check Compounds (Initial calibration)

System Performance check compounds (SPCCs)	Minimum average response factor (Method requirement)	Average Response factor 03-07-22
N-Nitroso-di-n-propylamine	≥ 0.5	√
Hexachlorocyclopentadiene	≥ 0.05	√
2,4-Dinitrophenol	≥ 0.01	√
4-Nitrophenol	≥ 0.01	√

√ denotes passing method acceptance limits

Average response factors for the rest of target compounds were within method's recommended values.

Calibration check compounds (CCCs) met the acceptance criteria for %RSD (less than 20%) among the response factors calculated for each level. Table 4.2.3.2 lists the CCCs with method requirement limits and calculated %RSD among response factors for initial calibration.

Table 4.2.3.2 Calibration Check Compounds (CCCs) Initial Calibration

Calibration Check Compounds (CCCs)	Response Factors %RSD (Method limit)	Accepted Response Factors 03-07-22
Phenol	≤ 20	√
1,4-Dichlorobenzene	≤ 20	√
2-Nitrophenol	≤ 20	√

Calibration Check Compounds (CCCs)	Response Factors %RSD (Method limit)	Accepted Response Factors 03-07-22
2,4-Dichlorophenol	≤ 20	√
Hexachlorobutadiene	≤ 20	√
4-Chloro-3-methylphenol	≤ 20	√
2,4,6-Trichlorophenol	≤ 20	√
Acenaphthene	≤ 20	√
N-Nitrosodiphenylamine	≤ 20	√
Pentachlorophenol	≤ 20	√
Fluoranthene	≤ 20	√
Di-n-Octylphthalate	≤ 20	√
Benzo(a)pyrene	≤ 20	√

√ denotes passing method acceptance limits

Average response factor curve fit was used mainly through the initial calibration. All target compounds met the maximum 15% RSD limit.

Least square linear regression was used for the following compounds where %RSD exceeded the maximum 15 percent limit.

Target Analytes	Least Square Linear Regression (CCF) 03-09-23
Benzoic acid	0.9969
2,4-Dinitrophenol	0.9964
4-Nitrophenol	0.9971
di-n-Octylphthalate	0.9972

All analytes met the acceptance criteria regarding minimum response factor and maximum %RSD.

4.2.4. Initial Calibration Verification and Continuing Calibration

Each set of initial calibration was verified by a second source standard on 03-07-22 and 03-31-23. Percent difference (%D) and/or drift between initial calibration RRFs (average response factors) and the initial calibration verification response factors (RFs) were less than or equal to 20% for all target compounds. Continuing calibration check standards were analyzed at the beginning and at the end of analysis shift on 04-28-23. Prior to continuing calibration injection, instrument performance tune check standard (DFTPP) was carried out. It passed all the method tuning criteria. Tailing of Pentachlorophenol and Benzidine was within acceptance limit (less than 2%) and breakdown of DDT was also within methods acceptance limit of less

than 20%. Minimum average response factors for the system performance check compounds (SPCCs) were all within the method limits according to the following table:

Table 4.2.4.1: System Performance Check Compounds (Daily calibration)

System Performance Check compounds (SPCCs)	Minimum response factor (Method limits)	Continuing cal. Response factors (04-28-23) I&II
N-Nitroso-di-n-propylamine	≥ 0.5	√
Hexachlorocyclopentadiene	≥ 0.05	√
2,4-Dinitrophenol	≥ 0.01	√
4-Nitrophenol	≥ 0.01	√

√ denotes passing method acceptance limits

Calculated percent differences (%D) between initial calibration RRFs (average response factors) and the continuing calibration response factors (RFs) were less than or equal to 20% for all the Calibration Check Compounds (CCCs) and less than or equal to 20% for nearly all other target analytes. Area counts for all internal standards were within ± 50 percent of the same level in the initial calibration. Percent difference between RFs from continuing calibration and average response factors from initial calibration passed the method's criteria as summarized in Table 4.2.4.2.

Table 4.2.4.2 Calibration Check Compounds (CCCs): Continuing Calibration

Calibration Check Compounds (CCCs)	% Deviation From Initial calibration (Acceptance Limit)	Accepted Deviation from Initial calibration (04-28-23) I & II
Phenol	≤ 20	√
1,4-Dichlorobenzene	≤ 20	√
2-Nitrophenol	≤ 20	√
2,4-Dichlorophenol	≤ 20	√
Hexachlorobutadiene	≤ 20	√
4-Chloro-3-methylphenol	≤ 20	√
2,4,6-Trichlorophenol	≤ 20	√
Acenaphthene	≤ 20	√
N-Nitrosodiphenylamine	≤ 20	√
Pentachlorophenol	≤ 20	√
Fluoranthene	≤ 20	√
Di-n-Octylphthalate	≤ 20	√
Benzo(a)pyrene	≤ 20	√

√ denotes passing method acceptance limits

Deviation from the initial calibration was less than 20 percent for the rest of target list, except for the compounds listed below in the closing continuing standard:

Target Analytes	%Deviation from Initial calibration Method Criteria	%Deviation from Initial calibration (04-28-23) I	%Deviation from Initial calibration (04-28-23) II
2,4-Dinitrophenol	≤ 20	--	26.7%*
Bis(2-Ethylhexylphthalate)	≤ 20	--	21.9%*
Caprolactam	≤ 20	--	22.9%*

*Exceeded 20% maximum allowed difference in the closing daily standard

This deviation should not affect quality of the presented data.

4.2.5. Quality Control samples reported consisted of one method blank, one set of LCS/LCSD and MS/MSD. Sample TMW63042023 was designated to be spiked as MS/MSD. The full list of target compounds was spiked and reported for LCS/LCSD and MS/MSD. Percent recoveries and percent RPDs for all the QC samples reported were all within the project acceptance limits, except for few compounds exceeding 20% maximum RPD limit in LCS/LCSD as listed in the table below:

Target analyte	LCS%	LCSD%	RPD >20%
Benzaldehyde	63	94	39%*
1,2,4,5-Tetrachlorobenzene	38	60	46%*
Hexachlorocyclopentadiene	41	52	24%*

*Outside QC limit

Results and recoveries of LCS/LCSD and MS/MSD was used to evaluate accuracy and precision for the rest of compound list in this method.

Results, percent recoveries and RPDs were recalculated randomly and all agreed with the reported QC summary table. Method blank presented with the data package, analyzed with samples did not show presence of any target compounds.

Surrogate recoveries were all within the method's acceptable limits except for two surrogates in one sample as shown in the table below:

Surrogate	TMW63042023 (D225-08)	QC Limit%
2-Fluorophenol	18%*	19-119
Phenol-D5	29%*	40-130

Surrogate	TMW63042023 (D225-08)	QC Limit%
Nitrobenzene d-5	42%*	44-120
2,4,6-Tribromophenol	35%*	43-140

*Outside QC limit

This was attributed to matrix interference in the case narrative.

4.2.6. Field duplicate sample and its associated sample: Field sample SMW01042023 was identified as field duplicate of sample SMW01042023D. No SVOCs was detected in any of field sample or associated field duplicate sample.

4.2.7. Raw data was submitted for all samples. Sample TMW50042023 (Lab ID# 23D225-03) was designated to be reviewed as stage 3 data deliverable. Raw data for this sample with the related QC samples was reviewed for stage 3 data validation. The results calculated from the raw data, agreed with all the results reported in data summary reports.

4.3. ORGANOCHLORINE PESTICIDES (EPA Method 3520C/ 8081B)

4.3.1. Technical Holding Times

Holding time from sample collection to extraction and extraction to analysis was met for seven water samples requested for this method. Water samples were collected on 04-19-23, extracted on 04-26-23 and analyzed on 05-02-23 and 05-03-23, within required holding time.

A dual column GC equipped with two Electron Capture Detectors (ECDs) was used for analysis. The results and raw data were submitted for both channel A and B. Channel A was used for data reporting.

DDT- Endrin breakdown mix was analyzed prior to initial calibration on 01-11-23, 01-09-23 and before sample analysis on 05-02-23, and 05-03-23. Breakdown of DDT to DDE and DDD and breakdown of Endrin-to-Endrin aldehyde and Endrin ketone were within the QC limits (less than 15 percent).

4.3.2. Initial Calibration

Initial calibration was performed with eight levels of concentration for each pesticide on 01-11-23. Both channels A and B were calibrated. Calibration factor (area for each compound/concentration) was used to quantify analytes. Average response factor was used to show linearity for both channels. Percent relative standard deviation (% RSDs) among calibration factors (CFs) for both channels A and B were within method specification (less than 15%). Pesticide target list was calibrated with two separate groups of compounds for each column. Due to interference with other pesticide target compounds, a separate curve was generated for Toxaphene on 01-09-23 for both channels. %RSD among the calibration factors was less than 15 for Toxaphene.

Retention time window width were established for all target analytes at each calibration level. Retention times for further sample analyses were within the assigned retention time windows set by initial calibration.

4.3.3. Initial Calibration Verification and Continuing Calibration

Initial calibration was verified by a second source standard (ICV) for all pesticide target list including Toxaphene, on 01-12-23 and 01-09-23. Percent difference between initial calibration response factors (Average response factors) and the response factors calculated for each analyte were less than 20% for both channels.

Performance of instrument was monitored by analysis of DDT and Endrin breakdown mixture. Before each continuing (daily) calibration a mixture of DDT and Endrin was analyzed. Breakdown of DDT to DDE and DDD and breakdown of Endrin to Endrin-aldehyde and Endrin-ketone were all less than 15%.

A total of three continuing calibration standards were analyzed at 10-injections interval on 05-02-23, and 05-03-23, bracketing the analyses of sample and all the QC samples. Percent difference between initial calibration average response factors and the response factors calculated for each analyte from continuing calibrations, were less than 20% for target compounds for both channels A and B. In all continuing calibration standards, one mid-point concentration of 20-40µg/L was injected. Results for surrogate recoveries and QC were all calculated from channel A. Channel B was used for confirmation only.

4.3.4. Quality Control samples consisted of method blank, one LCS and MS/MSD. Sample TMW63042023 was designated to be spiked for MS/MSD for this method. All pesticide target list was spiked and reported for LCS and MS/MSD. Percent recoveries (%R) were within established QC limits. Results and recoveries of QC samples were confirmed with the reported raw data.

Results for method blank were reviewed for each component and no organochlorine pesticide was found in the method blank.

Surrogate recoveries were all within the method QC acceptance limits.

4.3.5. Field duplicate sample and its associated sample: No duplicate sample was assigned to This method.

4.3.6. Raw data was submitted for all samples. Sample TMW50042023 (Lab ID# 23D225-03) was designated to be reviewed as stage 3 data deliverable. Raw data for this sample with the related QC samples was reviewed for stage 3 data validation. The results calculated from the raw data, agreed with all the results reported in data summary reports.

4.4. Polychlorinated Biphenyls (EPA Method 3520C/ 8082A)

4.4.1. Technical Holding Times

Holding time from sample collection to extraction and extraction to analysis was met for water samples requested for this method. Four water samples were collected on 04-19-23, extracted on 04-26-23, and analyzed on 04-27-23.

A dual column GC equipped with two Electron Capture Detectors (ECDs) was used for analysis. The results and raw data were submitted for both channel A and B. Channel A was used for data reporting.

4.4.2. Initial Calibration

Initial calibration was performed with seven concentration levels for Aroclor 1016 and Aroclor1260 on 01-24-23. Both channels A and B were calibrated. Calibration factor (area for each compound/concentration) was used to quantify analytes. Average response factor was used to show linearity for both channels. Percent relative standard deviation (% RSDs) among

calibration factors (CFs) for both channels A and B were within method specification (less than 20%).

Retention time window width were established at each calibration level. Retention times for further sample analyses were within the assigned retention time windows set by initial calibration.

4.4.3. Initial Calibration Verification and Continuing Calibration

Initial calibration was verified by a second source standard (ICV) for Aroclor 1016 and 1260, on 01-24-23. Percent difference between initial calibration response factors (Average response factors) and the response factors calculated for each analyte were less than 20% for both channels. After establishing linearity of the instrument through initial calibration, the rest of Aroclors, if required, were injected at single point for identification only.

Two continuing calibration standards were analyzed at 10-injections interval. It was carried out on 04-27-23, bracketing the analyses of sample and all the QC samples. Percent difference between initial calibration average response factors and the response factors calculated for each analyte from continuing calibrations, were less than 20% for each channel.

Results for surrogate recoveries and QC were all calculated from channel A. Channel B was used for confirmation only

4.4.4. Quality Control

samples consisted of one method blank, one LCS and MS/MSD.

Sample TMW63042023 was designated to be spiked as MS/MSD. Results and recoveries of LCS and MS/MSD was used to evaluate accuracy and precision. Percent recoveries (%R) were within the established QC limits.

Results for method blank was reviewed for each component and no target analyte was found in the method blank.

Surrogate recoveries were all within the method QC acceptance limits.

4.4.5. Field duplicate sample

and its associated sample: No duplicate sample was assigned to This method.

4.4.6. Raw data was submitted for all samples. Sample TMW50042023(Lab ID#23D225-03) with all related QC samples was reviewed for stage 3 data validation. The results calculated from the raw data, agreed with all the results reported in data summary reports.

4.5. Chlorinated Herbicides (EPA 8151A)

4.5.1. Technical Holding Times

Holding time from sample collection to extraction and extraction to analysis was met for water samples requested for this method. Four water samples were collected on 04-19-23, extracted on 04-24-23 and analyzed on 04-25-23 and 04-26-23.

A dual column GC equipped with two Electron Capture Detectors (ECDs) was used for analysis. The results and raw data were submitted for both channel A and B. Channel A was used for data reporting.

4.5.2. Initial Calibration

Initial calibration was performed with eight levels of concentration for each herbicide on 03-13-23. Both channels A and B were calibrated. Calibration factor (area for each compound/concentration) was used to quantify analytes. Average response factor was used to show linearity for both channels. Percent relative standard deviation (% RSDs) among calibration factors (CFs) for both channels A and B were within method specification (less than 20%) for all target list. Linear regression curve type with correlation coefficient of 0.9943 was used for MCP in column B.

Retention time windows were established for all target analytes at each calibration level. Retention times for further sample analyses were within the assigned retention time windows set by initial calibration.

4.5.3. Initial Calibration Verification and Continuing Calibration

Initial calibration was verified by a second source standard (ICV) for all target herbicides on 03-14-23. Percent difference between initial calibration response factors (Average response factors) and the response factors calculated for each analyte were less than 20% for both channels.

Two continuing calibration standards were analyzed at 10-injections interval. It was carried out on 04-25-23 and 04-26-23, bracketing the analyses of sample and all the QC samples. Percent difference between initial calibration average response factors and the response factors calculated for each analyte from continuing calibrations, were less than 20% for all target compounds in both channels.

Results for surrogate recoveries and QC were all reported from channel A. Channel B was used for confirmation only.

4.5.4. Quality Control samples consisted of one method blank, one set of LCS/LCSD and MS/MSD. Sample TMW63042023 was designated to be spiked as MS/MSD. The full herbicides target list was spiked and reported for LCS/LCSD and MS/MSD. Percent recoveries (%R) were within established QC limits

Results for method blank was reviewed for each component and no Herbicide was found in the method blank. Surrogate recoveries were all within the method QC acceptance limits.

4.5.5. Field duplicate sample and its associated sample: No field duplicate sample was assigned To this method.

4.5.6. Raw data was submitted for all samples. Sample TMW50042023 (Lab ID#23D225-03) with all related QC samples was reviewed for stage 3 data validation. The results calculated from the raw data, agreed with all the results reported in data summary reports.

4.6. Nitroaromatics by LC/MS/MS (EPA Method3535A/ 8330B)

4.6.1. Technical Holding Times

Holding time from sample collection to extraction and extraction to analysis was met for nine (9) water samples requested for this method. Water samples were collected on 04-19-23, extracted on 04-25-23 and analyzed on 05-22-23 and 05-23-23, within holding time.

A High-Performance LC (HPLC) coupled with Ultraviolet (UV) Detector was used for analysis. Positive results, if any, were confirmed with UHPLC equipped with different column (Kinetex- Biphenyl column).

4.6.2. Initial Calibration

Initial calibration was performed with seven concentration levels for each analyte on 08-25-22. Confirmation (Kinetex-Biphenyl column) was calibrated on 08-23-21. Calibration factor (area for each compound/concentration) was used to quantify analytes. Average response factor was used to show linearity for both channels. Percent relative standard deviation (% RSDs) among calibration factors (CFs) was within acceptable limit (less than 15 percent.)

Retention time windows were established for each target analyte at each calibration level. Retention times for further sample analyses were within the assigned retention time windows set by initial calibration.

4.6.3. Calibration Verification and Continuing Calibration

Initial calibration was verified by a second source standard (ICV) for each target analyte for both columns on 08-26-22 and 08-23-21. Percent difference between initial calibration response factors (Average response factors) and the response factors calculated for each analyte were less than 15% in both columns.

Continuing calibration standards were analyzed at 10-injections interval. A total of four continuing calibration standards were analyzed on 05-22-23, and 05-23-23, bracketing the analyses of samples and all the QC samples. Two more daily standards were analyzed with confirmation column on 05-30-23. Percent difference between initial calibration average response factors and the response factors calculated for each analyte from continuing calibrations were less than 15% for all analytes for both columns.

4.6.4. Quality Control samples consisted of one method blank, one set of LCS/LCSD and MS/MSD. Sample TMW63042023 was designated to be spiked as MS/MSD. The full explosive target lists were spiked and reported for LCS/LCSD and MS/MSD. Percent recoveries (%R) were within the QAPP established QC limits. Raw data for both un-spiked sample and spiked QC samples were matching the reported values.

Result for method blank was reviewed for each compound and no target was found in the method blank.

Surrogate recoveries were all within the method QC acceptance limits. 3,4-Dinitrotoluene was used as surrogate.

4.6.5 Field duplicate sample and its associated sample: Field sample SMW01042023 was identified as field duplicate of sample SMW01042023D. No explosives were detected in the field sample or associated field duplicate sample.

4.6.6. Raw data was submitted for all samples. Sample TMW50042023(Lab ID#23D225-03) was designated as stage 3 data deliverable. Raw data for this sample together with the related QC samples was reviewed for stage 3 data validation. The results calculated from the raw data, agreed with all the results reported in data summary reports. Positive results were confirmed by confirmation column.

4.7. Nitroglycerine and PETN by UHPLC/UV (EPA Method 8332)

4.7.1. Technical Holding Times

Holding time from sample collection to extraction and extraction to analysis was met for nine (9) water samples requested for this method. Water samples were collected on 04-19-23, extracted on 04-25-23 and analyzed on 05-15-23.

A High-Performance LC (HPLC) coupled with Ultraviolet (UV) Detector was used for analysis.

4.7.2. Initial Calibration

Initial calibration was performed with five levels of concentration for each analyte on 08-24-21. Calibration factor (area for each compound/concentration) was used to quantify analytes. Average response factor was used to show linearity. Percent relative standard deviation (% RSDs) among calibration factors (CFs) was within acceptable limit (less than 15 %.)

Retention time windows were established for each target analytes at each calibration level. Retention times for further sample analyses were within the assigned retention time windows set by initial calibration.

4.7.3. Calibration Verification and Continuing Calibration

Initial calibration was verified by a second source standard (ICV) for each analyte on 08-24-21. Percent difference between initial calibration response factors (Average response factors) and the response factors calculated for each analyte were less than 15%.

A total of three continuing calibration standards were analyzed at 10-injections interval. It was carried out on 05-15-23 and 05-16-23, bracketing the analyses of samples and all the QC samples. Percent difference between initial calibration average response factors and the response factors calculated for each analyte from continuing calibrations were less than 15% for each analyte.

4.7.4. Quality Control samples consisted of one method blank, one set of LCS/LCSD and MS/MSD. Sample TMW63042023 was designated to be spiked as MS/MSD. Each target compound was spiked and reported for LCS/LCSD and MS/MSD. Percent recoveries (%R) were within the established acceptance QC limits. Raw data for both un-spiked sample and spiked QC samples were matching the reported values.

Result for method blank was reviewed for each compound and no target was found in the method blank. Surrogate recoveries were all within the method QC acceptance limits.

4.7.5. Field duplicate sample and its associated sample: Sample SMW01042023 was identified as field duplicate of sample SMW01042023D. No explosive target compound was detected in sample or associated field duplicate sample.

4.7.6. Raw data was submitted for all samples. Sample TMW50042023 (Lab ID#23D225-03) was designated to be reviewed as stage 3 data deliverable. Raw data of this sample with the related QC samples was reviewed for stage 3 data validation. The results calculated from the raw data, agreed with all the results reported in data summary reports.

4.8. Total Petroleum hydrocarbons GRO (EPA Method 8015G)

4.8.1. Technical Holding Times

Holding time from sample collection to extraction and extraction to analysis was met for six water samples requested for this method. Water samples were collected on 04-19-23. Samples were analyzed on 04-20-23 and 04-21-23 within holding time requirement.

A GC coupled with Flame Ionization Detector (FID) was used for analysis. Sample was carried through the system by purge and trap.

4.8.2. Initial Calibration

Initial calibration was performed with six levels of concentration on 03-07-23. Calibration factor (area for each compound/concentration) was used to quantify gasoline range hydrocarbons (TPH as GRO). Average response factor was used to show linearity. Percent relative standard deviation (% RSDs) among calibration factors (CFs) was within acceptable limit (less than 15%.)

Retention time window width was established by analysis of window defining hydrocarbon standard (C6-C10). Retention times for further sample analyses was used for peak identification and integration range.

4.8.3. Initial Calibration Verification and Continuing Calibration

Initial calibration was verified by a second source standard (ICV) on 03-07-23. Percent difference between initial calibration response factors (Average response factors) and the response factors calculated for each analyte were less than 20%

Continuing calibration standards were analyzed at 10-injections interval. It was carried out on 04-20-23 and 04-21-23, bracketing the analyses of samples and all the QC samples. Percent difference between initial calibration average response factors and the response factors calculated for each group of GRO Hydrocarbons from continuing calibrations were less than 20%.

4.8.4. Quality Control samples consisted of one method blank, one set of LCS/LCSD and MS/MSD. Sample TMW63042023 was assigned to be spiked as MS/MSD. Percent recoveries (%R) were within the established QC limits for LCS/LCSD and MS/MSD. Raw data for both un-spiked sample and spiked QC samples were matching the reported values. Result for method blank was reviewed and no contamination was found in the method blank. Surrogate recoveries were all within the method QC acceptance limits.

4.8.5. Field duplicate sample and its associated sample: No field duplicate sample was assigned to this method.

4.8.6. Raw data was submitted for all samples. Sample TMW50042023(Lab ID#23D225-03) was designated as stage 3 data deliverable. Raw data for this sample with the related QC samples was

reviewed for stage 3 data validation. The results calculated from the raw data, agreed with all the results reported in data summary reports.

4.9. Total Petroleum hydrocarbons DRO (EPA Method 8015D)

4.9.1. Technical Holding Times

Holding time from sample collection to extraction and extraction to analysis was met for five water samples requested for this method. Water samples were collected on 04-19-23, extracted on 04-25-23 and analyzed on 04-26-23, within holding time.

A GC coupled with Flame Ionization Detector (FID) was used for analysis. Heavier range of total petroleum hydrocarbons were extracted and introduced into system by direct injection.

4.9.2. Initial Calibration

Initial calibration was performed with seven levels of concentration on 02-17-23. Calibration factor (area for each compound/concentration) was used to quantify diesel range hydrocarbons (TPH as DRO). A second set of initial calibration curve was generated for lighter TPHs (Jet Fuel) and heavier TPHs (motor oil). Average response factor was used to show linearity. Percent relative standard deviation (% RSDs) among calibration factors (CFs) was within acceptable limit (less than 15 %.)

Retention time window width was established by analysis of a window defining hydrocarbon standard (C10-C40). Retention times for further sample analyses was used for peak identification and integration range.

4.9.3. Initial Calibration Verification and Continuing Calibration

Initial calibration was verified by a second source standard (ICV) on 02-17-23. Percent difference between initial calibration response factors (Average response factors) and the response factors calculated for each analyte were less than 20%

Three continuing calibration standards were analyzed at 10-injections interval. It was carried out on 04-26-23 bracketing the analyses of samples and all the QC samples. Percent difference between initial calibration average response factors and the response factors calculated for each DRO group from continuing calibrations were less than 20%.

4.9.4. Quality Control samples consisted of one method blank, one set of LCS/LCSD and MS/MSD. Sample TMW63042023 was designated to be spiked as MS/MSD. Percent recoveries (%R) of LCS/LCSD and MS/MSD were within the QAPP established QC limits. Raw data for both un-spiked sample and spiked QC samples were matching the reported values.

Result for method blank was reviewed and no contamination was found in the method blank. Surrogate recoveries were all within the method QC acceptance limits.

4.9.5. Field duplicate sample and its associated sample: No field duplicate sample was assigned to this method.

4.9.6. Raw data was submitted for all samples. Sample TMW50042023(Lab ID#23D225-03) was designated to be reviewed as stage 3 data deliverable. Raw data for this sample with the related QC samples was reviewed for stage 3 data validation. The results calculated from the raw data, agreed with all the results reported in data summary reports.

4.10. Perchlorate by HPLC/MS/MS (EPA Method 6850)

4.10.1. Technical Holding Times

Holding time from sample collection to extraction and extraction to analysis was met for eleven samples requested for this method. Water samples were collected on 04-19-23. Samples were analyzed on 04-28-23, 04-29-23, 05-01-23 and 05-03-23, within holding time.

A High-Performance LC coupled with Mass Detector (HPLC/MS) was used for analysis.

4.10.2. Initial Calibration

Initial calibration was performed with seven levels of concentration on 03-13-23. Internal standard curve type was used for quantifying Perchlorate. Isotopically-labeled Perchlorate ion ($\text{Cl}^{18}\text{O}_4^-$) was added to serve both as internal standard and correction for Perchlorate loss from sample preparation. The correlation coefficient of 0.9962 (Perchlorate ion 83) and 0.9986 (perchlorate ion 85) was calculated to show the linearity of each curve. The concentrations used for calibration ranged from 0.1 – 7.5 µg/L.

Retention time for each isotope (ion 83 and 85) at each calibration level was within 0.2 minutes required by the method.

4.10.3. Initial Calibration Verification and Continuing Calibration

Initial calibration was verified by a second source standard (ICV) on 03-13-23. Percent recoveries were within required method limits (85-115% of the true value).

Continuing calibration standards were analyzed at 10-injections interval. A total of fourteen daily standards were carried out on 04-28-23, 04-29-23, 05-01-23 and 05-03-23, bracketing the analyses and re-analysis of samples and all the QC samples. Recoveries of continuing calibration standards were within 90-110% limit.

4.10.4. Quality Control samples consisted of three method blanks, three sets of LCS/LCSD and MS/MSD. Sample TMW63042023 was designated to be spiked as MS/MSD. Perchlorate was spiked and reported for LCS/LCSD and MS/MSD. Percent recoveries (%R) were within the established QC limits. Raw data for both un-spiked sample and spiked QC samples were matching the reported values.

Result for method blanks was reviewed for each compound and no target was found in the method blank. No surrogate is used in this method.

4.10.5. Field duplicate sample and its associated sample: Field sample SMW01042023 was identified as field duplicate of sample SMW01042023D. No Perchlorate was detected in the field sample or associated field duplicate sample.

4.10.6. Raw data was submitted for all samples. Sample TMW50042023 (Lab ID#23D225-03) was designated to be reviewed as stage 3 data deliverable. Raw data for this sample with the related QC samples was reviewed for stage 3 data validation. The results calculated from the raw data, agreed with all the results reported in data summary reports.

4.11. TOTAL and DISSOLVED METALS BY ICP (EPA Method 6020A)

4.11.1. Technical Holding Times

Holding time from sample collection to extraction and extraction to analysis was met for eleven water samples requested for this method. Water samples were collected on 04-19-23. Samples were prepared (digested) on 05-02-23 for both total and dissolved metals. Samples and QC samples were analyzed on 06-28-23 for total and dissolved metals by ICP MS. Water

samples were preserved and filtered in the lab for dissolved metals analysis. Therefore, two sets of data were generated, one for total metals and one for dissolved metals.

4.11.2. Initial Calibration and Continuing calibration standards

Initial calibration was performed at the start of analysis day on 06-28-23. Tune performance report was generated at the start of each analysis day, before initial calibration. It was within method's acceptance criteria. Initial instrument (ICP-MS) calibration for this method was acceptable. Five levels of concentration were used for each initial calibration. One method blank and one calibration standard was used for each daily check standard. A range of concentrations of standards was used for calibration. The concentrations used are summarized as follow:

Metals	Concentration µg/L
Se, Pb, Sb, As, Ba, Be, Cd, Cr, Co, Cu, Mn, Ni, Ag, Tl, V, and Zn	50,100,500,1000
Al, Fe, Ca, Mg, Na, and K	50,000

Initial and continuing calibration verification standards for each element was within acceptable limit of 90-110 percent of the true value. Continuing calibration standards were analyzed at the frequency required by the method. Results for the ICP Interference Check Solutions (ICS-A and ICS-AB) were within the control limits of $\pm 20\%$ of the true value for the analytes incorporated in each solution.

4.11.3. Quality Control samples consisted of method blank, one set of LCS/LCSD and MS/MSD for total and dissolved metals. Sample TMW63042023 was designated to be spiked as MS/MSD. Recoveries of LCS/LCSD and MS/MSD were all within the acceptance for both total and dissolved metals. However, few metals failed the acceptable QC limits in MS/MSD as shown in the table below:

ANALYTE	Total Metals			Dissolved Metals		
	TMW63042023 MS%	TMW63042023 MSD%	QC Limit %	TMW63042023 MS%	TMW63042023 MSD%	QC Limit %
Aluminum	√	√	84-117	√	√	84-117
Antimony	√	√	85-117	√	√	85-117
Arsenic	√	√	84-116	√	√	84-116
Barium	√	√	86-114	√	√	86-114
Beryllium	√	√	83-121	√	√	83-121
Cadmium	√	√	87-115	√	√	87-115
Calcium	114*	134*	87-118	121*	111*	87-118
Chromium	√	√	85-116	√	√	85-116
Cobalt	√	√	86-115	√	√	86-115
Copper	√	√	85-118	√	√	85-118
Iron	√	√	87-118	√	√	87-118
Lead	√	√	88-115	√	√	88-115
Magnesium	√	√	83-118	√	√	83-118
Manganese	95	126*	87-115	√	√	87-115
Nickel	√	√	85-117	√	√	85-117
Potassium	√	√	85-115	√	√	85-115
Selenium	√	√	80-120	√	√	80-120
Silver	√	√	85-116	√	√	85-116
Sodium	700*	133*	85-117	400*	900*	85-117
Thallium	√	√	82-116	√	√	82-116
Vanadium	√	√	86-115	√	√	86-115
Zinc	√	√	83-119	76*	72*	83-119

*Outside control limits

Results for method blanks were acceptable and no contamination was found in the method blanks. Calibration blanks were analyzed after each continuing calibration standard. Sample TMW63042023 was also used for serial dilution for both total and dissolved metals. The serial dilution analysis (at 5-fold dilution) was within 10% difference of the initial analysis, except for Manganese; (total metals at 28% and dissolved metals at 37%). The same sample was used for spike addition (analytical spike). All results were within the QC limit of (80-120%) for total and dissolved metals.

4.11.4. Field duplicate sample and its associated sample: Field sample SMW01042023 was identified as field duplicate of sample SMW01042023D. Results and %RPD for field sample and associated field duplicate sample are listed in the table below:

ANALYTE	Total Metals			Dissolved Metals		
	SMW01042023 µg/L	SMW01042023D µg/L	% RPD	SMW01042023 µg/L	SMW01042023D µg/L	% RPD
Aluminum	U	U	--	U	U	--
Antimony	U	U	--	U	U	--
Arsenic	2.7J	2.6J	3.77	2.3J	2.4J	--
Barium	36	36	<1	36	34	5.71
Beryllium	U	U	--	U	U	--
Cadmium	U	U	--	U	U	--
Calcium	20000	19000	5.13	19000	19000	<1
Chromium	U	U	--	U	U	--
Cobalt	U	U	--	U	U	--
Copper	U	U	--	U	U	--
Iron	U	U	--	U	U	--
Lead	U	U	--	U	U	--
Magnesium	6600	6600	<1	6600	6400	3.1
Manganese	270	270	<1	290	270	7.14
Nickel	1.9J	1.9J	<1	2.2J	1.9J	14.6
Potassium	U	U	--	U	U	--
Selenium	U	U	--	U	U	--
Silver	U	U	--	U	U	--
Sodium	1100000	1100000	<1	1100000	1100000	<1
Thallium	U	U	--	U	U	--
Vanadium	6.1	6.3J	3.22	6.1	5.9	3.33
Zinc	48J	47J	2.10	30J	29J	3.39

4.11.5. Raw data was submitted for all samples. Sample TMW50042023 (Lab ID#23D225-03) was designated to be reviewed as stage 3 data deliverable. Raw data for this sample with the related QC samples was reviewed for stage 3 data validation. The results calculated from the raw data, agreed with all the results reported in data summary reports.

4.12. MERCURY & Dissolved Mercury by COLD VAPOR: EPA Method 7470A

4.12.1. Technical Holding Times

Holding time from sample collection to extraction and extraction to analysis was met for eleven (11) water samples requested for this method. Water samples were collected on 04-19-23. Samples were prepared (digested) on 05-04-23 for Mercury and dissolved Mercury. Samples were analyzed for Mercury and dissolved Mercury on 05-04-23. All samples were preserved and filtered in the lab for dissolved Mercury analysis.

4.12.2. Initial and continuing calibrations: Instrument calibration for this method were acceptable. One blank and five standard levels were used for calibration curve at the beginning of analysis day on 05-04-23. The correlation coefficient of at least 0.999419 was calculated to show the linearity of calibration curve. The concentrations used for calibration ranged from 0.2 – 5.0 µg/L.

Initial calibration verification and Continuing calibration verification standards were within the acceptable range (90-110% of the spiked value).

4.12.3. Quality Control: Data for all the QC samples were within acceptable control limits. The QC samples consisted of one method blank, one set of LCS/LCSD and MS/MSD for Mercury and dissolved Mercury. Sample TMW63042023 was designated to be spiked as MS/MSD. Percent recoveries and % differences were within the control limits for LCS/LCSD and MS/MSD for both Mercury and dissolved Mercury. Sample TMW63042023 was also used for serial dilution. The results were all within QC acceptable limits. Method blank data was reviewed and no Mercury contamination was found in method blanks.

4.12.4. Field duplicate sample and its associated sample: Field sample SMW01042023 was identified as field duplicate of sample SMW01042023D. No Mercury was detected in sample and associated field duplicate sample.

4.12.5. Raw data was submitted for all samples. Sample TMW50042023 (Lab ID#23D225-03) was designated to be reviewed as stage 3 data deliverable. Raw data for this sample with the related QC samples was reviewed for stage 3 data validation. The results calculated from the raw data, agreed with all the results reported in data summary reports.

4.13. Method SW9056A: Nitrate-N, Nitrite, Orthophosphate, Bromide, Fluoride Sulfate and Chloride

4.13.1. Technical Holding Times: Holding time from sample collection to analysis was met for analysis of ten water samples requested for this method. Water samples were collected on 04-19-23. Samples were analyzed on 04-19-23 and 04-20-23, within the required 48-hour holding time for Nitrate, Nitrite, Orthophosphate, Bromide and Fluoride. Samples were re-

analyzed at higher dilutions on 04-21-23, 04-22-23 and 04-23-23 for Chloride and Sulfate within 28-day holding times.

4.13.2. Initial and continuing calibration: Anions such as Chloride, Fluoride, Bromide, Nitrite, Nitrate, Orthophosphate and Sulfate were separated from water samples by Ion chromatography. The separated anions in their acid form (very conductive) were measured by conductivity. They were identified on the basis of retention time as compared to reference standards.

Instrument was initially calibrated with nine calibration levels (0.05-20mg/L) on 11-09-22. Linear curve type with correlation coefficient of at least 0.999 for each anion was used throughout analysis. Percent RSD among calibration factors was less than 15%. Calibration curve (concentration versus area count of each anion) was presented for each anion. Area for each level was randomly checked with the values used in each calibration curve. All agreed with the raw data. A second source standard mixture (ICV) was used to verify the linearity of initial calibration on 11-09-22. Recoveries were all within 90-110% of initial true value. Continuing Calibration standards at 10-injections interval were analyzed on 04-19-23, 04-20-23, 04-21-23, and 04-22-23. A total of eleven continuing calibration standards were analyzed with analysis and re-analysis of all samples. In all continuing calibration standards submitted, the recoveries of target anions were within 90-110% of the expected values. After each continuing calibration standard, one calibration blank was injected. All the blanks were reported as non-detected regarding all the target anions. Retention time window width was established and confirmed with reference standard. It was within the assigned QC limit for each anion.

4.13.3. Quality Control Samples consisted of two method blanks, two sets of LCS/LCSD, MS/MSD and sample/sample duplicate. Sample TMW63042023 was designated to be spiked as MS/MSD. The same sample was analyzed as sample/sample duplicate for all anions. Recoveries of LCS/LCSD and MS/MSD were all within 90-110 % of spiked values for each anion. Percent RPDs for each sample/sample duplicate analysis was within required acceptance limit. However, Bromide failed the lower acceptance limit in MS/MSD as indicated in the table below:

Anion	Total Metals		
	TMW63042023 MS%	TMW63042023 MSD%	QC Limit%
Bromide	85*	85*	91-110

. Percent RPDs were less than 20% for LCS/LCSD and MS/MSD recoveries.

4.13.4. Field duplicate sample and its associated sample: Field sample SMW01042023 was identified as field duplicate of sample SMW01042023D. Results for sample/sample duplicates are summarized in the table below:

Anion	SMW01042023 (Lab ID #D225-04) mg/L	SMW01042023D (Lab ID #D225-05) mg/L	% RPD
Nitrate	0.45J	U	200
Nitrite	U	U	--
Orthophosphate	U	U	--
Bromide	1.1	1.1	<1
Chloride	660	670	1.5
Fluoride	1.5	1.5	<1
Sulfate	730	750	2.70

4.13.5. Raw data was submitted for all requested field samples. Sample TMW50042023 (Lab ID#23D225-03) was designated to be reviewed as stage 3 data deliverable. Raw data for this sample together with all related QC samples was reviewed for stage 3 data validation. All the samples were analyzed according to the prescribed QC procedures. All criteria were met. Due to high concentrations of Sulfate and Chloride, samples were analyzed with high dilution factors on 04-21-23, 04-22-23 and 04-23-23 within 28-day holding time.

5.0 CONCLUSION

SDG #23D225 analytical data evaluated in this data validation report has met the data quality and usability requirement as defined in the data quality objectives. The qualified QC data, if any, was due to matrix interference in the parent sample. Overall analytical data is of acceptable quality and considered usable for its intended purpose.

6.0 REFERENCES

1. *USEPA Analytical Operations/Data Quality Center (AOC) National Functional Guidelines for Organic Data Review* (USEPA, January 2017).
2. *USEPA Analytical Operations/Data Quality Center (AOC) National Functional Guidelines for Inorganic Data Review* (USEPA, September 2016).
3. *Quality Assurance Project Plan, final Draft, USACE Fort Wingate, NM* (Project No. Eco-18-1237, Eco & Associates Inc. April 2019).
4. U.S. Environmental Protection Agency, Dec. 1996, *SW846 Laboratory Manual Physical/Chemical Methods*. Revision 3, Washington, D.C. 20460.
5. EM 200-1-10 Guidance for Evaluation Performance-based Chemical Data, US Army Corps of Engineers (USACE), June 2005.
6. Department of Defense Quality System Manual (DOD QSM), Version 5.3, 2019
7. EPA Methods for Chemical Analysis of Water and Wastes. EPA -600-4-79-020. Revised; March 1983.